




SOLAR CELL I-V TEST SYSTEM USER MANUAL

Manual version: 3.0.7
Product code: T2002/T2003
Product Version: 3.0
Software version: 2.2



Contents

1. Overview	3
2. EU Declaration of Conformity (DoC).....	4
3. Safety.....	7
3.1 Use of Equipment.....	7
3.2 Hazard Icons	7
3.3 General Hazards	7
3.4 Power Cord Safety	8
3.5 Servicing	8
3.6 Health and Safety – Servicing	8
4. Requirements.....	9
5. Unpacking.....	9
5.1 Packing List.....	9
5.2 Damage Inspection.....	9
6. Specifications.....	10
7. System Components.....	11
8. Installation.....	12
9. Operation.....	12
9.1 Measurement Types.....	12
9.2 Quick Start Guide.....	14
9.3 Shared Software Settings.....	15
9.4 Characterisation Settings.....	17
9.5 Lifetime Settings	18
9.6 Stabilised Current Settings	20
9.7 Saving and Loading Settings.....	22
9.8 Saving Results	22
9.9 Controls	24
9.10 Plot Controls.....	25
10. Test Devices	26
10.1 Taking a Measurement	26
11. Troubleshooting	29
11.1 Installation and Setup	29
11.2 Error Messages and Warnings	30
12. Related Products	31
12.1 Related Consumables	31
12.2 Related Equipment.....	31

1. Overview

The Ossila Solar Cell I-V Test System is a low-cost solution for reliable current-voltage characterisation of solar cells. The system is controlled by specially designed software which can perform multiple I-V measurements, determine key metrics of solar cells, and measure these properties over long periods of time. The automated version of the system enables automatic switching between pixels, making measurements faster and easier.



2. EU Declaration of Conformity (DoC)

We

Company Name: Ossila BV

Postal Address: Biopartner 3 building, Galileiweg 8

Postcode: 2333 BD Leiden

Country: The Netherlands

Telephone number: +31 (0)718 081020

Email Address: info@ossila.com

declare that the DoC is issued under our sole responsibility and belongs to the following product:

Product: Solar Cell I-V Test System – Manual

(T2002A2/T2002B2/T2002D2/T2002E2/T2002F2/T2002G2), Cell I-V Test System – Automated (T2003A3/T2003B3/T2003C3/T2003E3/T2003F3/T2003G3)

Serial number: T2002A2-xxxx, T2002B2-xxxx, T2002D2-xxxx, T2002E2-xxxx, T2002F2-xxxx, T2003A3-xxxx, T2003B3-xxxx, T2003C3-xxxx, T2003E3-xxxx, T2003F3-xxxx, T2003G3.

Object of declaration:

Solar Cell I-V Test System – Manual (T2002A2/T2002B2/T2002D2/T2002E2/T2002F2/T2002G2)

Solar Cell I-V Test System – Automated (T2003A3/T2003B3/T2003C3/T2003E3/T2003F3/T2003G3)

The object of declaration described above is in conformity with the relevant Union harmonisation legislation:

EMC Directive 2014/30/EU

RoHS Directive 2011/65/EU

Signed:



Name: Dr James Kingsley

Place: Leiden

Date: 16/11/2021

Декларация за съответствие на ЕС

Производител: Ossila BV, Biopartner 3 building, Galileiweg 8, 2333 BD Leiden, NL.

Декларира с цялата си отговорност, че посоченото оборудване съответства на приложимото законодателство на ЕС за хармонизиране, посочено на предходната(-ите) страница(-и) на настоящия документ.

[Čeština] Prohlášení o shodě EU

Výrobce: Ossila BV, Biopartner 3 building, Galileiweg 8, 2333 BD Leiden, NL.

Prohlašujeme na vlastní odpovědnost, že uvedené zařízení je v souladu s příslušnými harmonizačními předpisy EU uvedenými na předchozích stranách tohoto dokumentu.

[Dansk] EU-overensstemme Iseserklæring

Producent: Ossila BV, Biopartner 3 building, Galileiweg 8, 2333 BD Leiden, NL.

Erklærer herved, at vi alene er ansvarlige for, at det nævnte udstyr er i overensstemmelse med den relevante EU-harmoniseringslovgivning, der er anført på den/de foregående side(r) i dette dokument.

[Deutsch] EU-Konformitätserklärung

Hersteller: Ossila BV, Biopartner 3 building, Galileiweg 8, 2333 BD Leiden, NL.

Wir erklären in alleiniger Verantwortung, dass das aufgeführte Gerät konform mit der relevanten EU-Harmonisierungsgesetzgebung auf den vorangegangenen Seiten dieses Dokuments ist.

[Eesti keel] ELi vastavusavaldus

Tootja: Ossila BV, Biopartner 3 building, Galileiweg 8, 2333 BD Leiden, NL.

Kinnitame oma ainuvastutust, et loetletud seadmed on kooskõlas antud dokumendi eelmisel leheküljel / eelmistel lehekülgedel ära toodud asjaomaste ELi ühtlustamise õigusaktidega.

[Ελληνικά] Δήλωση πιστότητας ΕΕ

Κατασκευαστής: Ossila BV, Biopartner 3 building, Galileiweg 8, 2333 BD Leiden, NL.

Δηλώνουμε υπεύθυνα ότι ο αναφερόμενος εξοπλισμός συμμορφώνεται με τη σχετική νομοθεσία εναρμόνισης της ΕΕ που υπάρχει στις προηγούμενες σελίδες του παρόντος εγγράφου.

[Español] Declaración de conformidad UE

Fabricante: Ossila BV, Biopartner 3 building, Galileiweg 8, 2333 BD Leiden, NL.

Declaramos bajo nuestra única responsabilidad que el siguiente producto se ajusta a la pertinente legislación de armonización de la UE enumerada en las páginas anteriores de este documento.

[Français] Déclaration de conformité UE

Fabricant: Ossila BV, Biopartner 3 building, Galileiweg 8, 2333 BD Leiden, NL.

Déclarons sous notre seule responsabilité que le matériel mentionné est conforme à la législation en vigueur de l'UE présentée sur la/les page(s) précédente(s) de ce document.

[Hrvatski] E.U izjava o sukladnosti

Proizvođač: Ossila BV, Biopartner 3 building, Galileiweg 8, 2333 BD Leiden, NL.

Izjavljujemo na vlastitu odgovornost da je navedena oprema sukladna s mjerodavnim zakonodavstvom EU-a o usklađivanju koje je navedeno na prethodnoj(nim) stranici(ama) ovoga dokumenta.

[Italiano] Dichiarazione di conformità UE

Produttore: Ossila BV, Biopartner 3 building, Galileiweg 8, 2333 BD Leiden, NL.

Si dichiara sotto la propria personale responsabilità che l'apparecchiatura in elenco è conforme alla normativa di armonizzazione UE rilevante indicata nelle pagine precedenti del presente documento.

[Latviešu] ES atbilstības deklarācija

Ražotājs: Ossila BV, Biopartner 3 building, Galileiweg 8, 2333 BD Leiden, NL.

Ar pilnu atbildību paziņojam, ka uzskaitītais aprīkojums atbilst attiecīgajiem ES saskaņošanas tiesību aktiem, kas minēti iepriekšējās šī dokumenta lapās.

[Lietuvių k.] ES atitikties deklaracija

Gamintojas: Ossila BV, Biopartner 3 building, Galileiweg 8, 2333 BD Leiden, NL.

atsakingai pareiškia, kad išvardinta įranga atitinka aktualius ES harmonizavimo teisės aktus, nurodytus ankstesniuose šio dokumento

[Magyar] EU-s megfelelőségi nyilatkozat

Gyártó: Ossila BV, Biopartner 3 building, Galileiweg 8, 2333 BD Leiden, NL.

Kizárólagos felelősségünk mellett kijelentjük, hogy a felsorolt eszköz megfelel az ezen dokumentum előző oldalán/oldalain található EU-s összehangolt jogszabályok vonatkozó rendelkezéseinek.

[Nederlands] EU-Conformiteitsverklaring

Fabrikant: Ossila BV, Biopartner 3 building, Galileiweg 8, 2333 BD Leiden, NL.

Verklaart onder onze uitsluitende verantwoordelijkheid dat de vermelde apparatuur in overeenstemming is met de relevante harmonisatiewetgeving van de EU op de vorige pagina('s) van dit document.

[Norsk] EU-samsvarserklæring

Produsent: Ossila BV, Biopartner 3 building, Galileiweg 8, 2333 BD Leiden, NL.

Erklærer under vårt eneansvar at utstyret oppført er i overholdelse med relevant EU-harmoniseringslovverk som står på de(n) forrige siden(e) i dette dokumentet.

[Polski] Deklaracja zgodności Unii Europejskiej

Producent: Ossila BV, Biopartner 3 building, Galileiweg 8, 2333 BD Leiden, NL.

Oświadczamy na własną odpowiedzialność, że podane urządzenie jest zgodne ze stosownymi przepisami harmonizacyjnymi Unii Europejskiej, które przedstawiono na poprzednich stronach niniejszego dokumentu.

[Por tuguês] Declaração de Conformidade UE

Fabricante: Ossila BV, Biopartner 3 building, Galileiweg 8, 2333 BD Leiden, NL.

Declara sob sua exclusiva responsabilidade que o equipamento indicado está em conformidade com a legislação de harmonização relevante da UE mencionada na(s) página(s) anterior(es) deste documento.

[Română] Declarație de conformitate UE

Producător: Ossila BV, Biopartner 3 building, Galileiweg 8, 2333 BD Leiden, NL.

Declară pe proprie răspundere că echipamentul prezentat este în conformitate cu prevederile legislației UE de armonizare aplicabile prezentate la pagina/paginile anterioare a/ale acestui document.

[Slovensky] Vyhlásenie o zhode pre EÚ

Výrobca: Ossila BV, Biopartner 3 building, Galileiweg 8, 2333 BD Leiden, NL.

Na vlastnú zodpovednosť prehlasuje, že uvedené zariadenie je v súlade s príslušnými právnymi predpismi EÚ o harmonizácii uvedenými na predchádzajúcich stranách tohto dokumentu.

[Slovenščina] Izjava EU o skladnosti

Proizvajalec: Ossila BV, Biopartner 3 building, Galileiweg 8, 2333 BD Leiden, NL.

s polno odgovornostjo izjavlja, da je navedena oprema skladna z veljavno uskladitveno zakonodajo EU, navedeno na prejšnji strani/prejšnjih straneh tega dokumenta.

[Suomi] EU-vaatimusten mukaisuusvakuutus

Valmistaja: Ossila BV, Biopartner 3 building, Galileiweg 8, 2333 BD Leiden, NL.

Vakuutamme täten olevamme yksin vastuussa siitä, että tässä asiakirjassa luetellut laitteet ovat tämän asiakirjan sivuilla edellisillä sivuilla kuvattujen olennaisten yhdenmukaistamista koskevien EU-säädösten vaatimusten mukaisia.

[Svenska] EU-försäkran om överensstämmelse

Tillverkare: Ossila BV, Biopartner 3 building, Galileiweg 8, 2333 BD Leiden, NL.

Vi intygar härmed att den utrustning som förtecknas överensstämmer med relevanta förordningar gällande EU-harmonisering som finns på föregående sidor i detta dokument.

3. Safety

3.1 Warning

- Do NOT connect external voltage sources to either SMU channel.
- The absolute maximum input voltage for the Vsense channels is ± 12 V. Do NOT apply input while not powered.

3.2 Use of Equipment

The Ossila Solar Cell I-V Test System is designed to be used as instructed. It is intended for use under the following conditions:

- Indoors in a laboratory environment (Pollution Degree 2).
- Altitudes up to 2000m.
- Temperatures of 5°C to 40°C; maximum relative humidity of 80% up to 31°C.

The unit is supplied with a 24 VDC power adapter with a power cord for the country of purchase, in accordance with European Commission regulations and British Standards. Use of any other electrical power cables, adaptors, or transformers is not recommended

3.3 Hazard Icons

The following symbols can be found at points throughout the rest of the manual. Note and read each warning before attempting any associated operations associated with it:

Table 3.1. Hazard warning labels used in this manual.

Symbol	Associated Hazard
	Electrical shock

3.4 General Hazards

Before installing or operating the Ossila Solar Cell I-V Test System there are several health and safety precautions which must be followed and executed to ensure safe installation and operation.

3.5 Power Cord Safety



Emergency power disconnect options: use the power cord as a disconnecting method and remove from wall. To facilitate disconnect, make sure the power outlet for this cord is readily accessible to the operator.

3.6 Servicing

If servicing is required, please return the unit to Ossila Ltd. The warranty will be invalidated if:

- Modification or service has taken place by anyone other than an Ossila engineer.
- The Unit has been subjected to chemical damage through improper use.
- The Unit has been operated outside the usage parameters stated in the user documentation associated with the Unit.
- The Unit has been rendered inoperable through accident, misuse, contamination, improper maintenance, modification, or other external causes.

3.7 Health and Safety – Servicing



Servicing should only be performed by an Ossila engineer. Any modification or alteration may damage the equipment, cause injury, or death. It will also void your equipment's warranty.

4. Requirements

Table 4.1 details the power requirements for the Solar Cell I-V Test System, and the minimum computer specifications for the Ossila Solar Cell I-V software.

Table 4.1. Solar Cell I-V Test System requirements.

Power	24 VDC
Operating System	Windows 10 or 11 (64-bit)
CPU	Dual Core 2 GHz
RAM	4 GB
Available Hard Drive Space	241 MB
Monitor Resolution	1680 x 1050
Connectivity	USB 2.0 Ethernet (requires DHCP)

5. Unpacking

5.1 Packing List

The standard items included with the Ossila Solar Cell I-V Test System are:

- Ossila Solar Cell I-V Test System.
- 24 VDC power adapter.
- USB-B cable.
- USB memory stick pre-loaded with the user manual, USB drivers, quality control data, and software installer.
- Resistor test device.

5.2 Damage Inspection

Examine the components for evidence of shipping damage. If damage has occurred, please contact Ossila directly for further action. The shipping packaging will come with a shock indicator to show if there has been any mishandling of the package during transportation.

6. Specifications

The Solar Cell I-V Test System specifications are shown in **Table 6.1**.

Table 6.1. Ossila Solar Cell I-V Test System specifications.

Voltage range	± 10 mV to ± 10 V
Voltage accuracy	± 10 mV offset
Voltage resolution	170 μ V
Current range	± 10 nA to ± 200 mA (5 ranges)
Current accuracy	± 10 nA (at 20 μ A range)
Current resolution	0.1 nA (at 20 μ A range)
Substrate Size	20 mm x 15 mm 25 mm x 25 mm 75 mm x 25 mm
Substrate Compatibility	T2002B, T2003B – S211 T2002E, T2003E – S2006 T2002F, T2003F – S241, S251
Overall Dimensions (Automated)	Width: 150 mm Height: 55 mm Depth: 300 mm
Overall Dimensions (Manual)	Source Measure Unit Width: 125 mm; Height: 55 mm; Depth: 185 mm Test Board Width: 105 mm; Height: 40 mm; Depth: 125 mm (T2002F - Width: 100 mm; Height: 40 mm; Depth: 150 mm)

7. System Components

The Solar Cell I-V Test System is comprised of 2 items: the Solar Cell I-V Test System (Figure 7.1 or Figure 7.2) and the Ossila I-V Curve software (Figure 7.3).



Figure 7.1. Solar Cell I-V Test System (Automated).




Figure 7.2. Solar Cell I-V Test System (Manual): a Source Measure Unit and Push-Fit Test Board.




Figure 7.3. Solar Cell I-V Test System software.

8. Installation

1. Install the Ossila Solar Cell I-V software on your PC.
 - I. Run the file 'Ossila-Solar-Cell-IV-Installer-vX-X-X-X.exe' on the USB memory stick provided.
 - II. Follow the on-screen instructions to install the software.
2. Connect the 24 VDC power adaptor to the power socket on the rear of the unit.
3. Connect the unit to your PC using the provided USB-B cable, or an Ethernet cable if preferred.

Note: The Ossila Solar Cell I-V software can also be downloaded from www.ossila.com/pages/software-drivers

9. Operation

9.1 Measurement Types

The Solar Cell I-V software can perform 3 different types of measurements. Each measurement type can be selected using the tabs at the top of the window. The available measurements are:

1. Characterisation (**Section 9.1.1**).
2. Lifetime (**Section 9.1.2**).
3. Stabilised Current (**Section 9.1.3**).

Each measurement type requires several settings to be selected before it can be performed. Settings that are shared between all measurements are detailed in **Section 9.3**. Measurement-specific settings are detailed in **Sections 9.4, 9.5, and 9.6**.

9.1.1 Characterisation

The Characterisation tab performs current-voltage (I-V) measurement and analysis of solar cells. The analysis calculates the following properties:

- Power conversion efficiency (PCE)
- Fill factor (FF)
- Short-circuit current density (J_{sc})
- Open-circuit voltage (V_{oc})
- Shunt resistance (R_{sh})
- Series resistance (R_s)
- Maximum power (P_{max})




Figure 9.1. Solar Cell I-V software: The Characterisation tab.

9.1.2 Lifetime

The Lifetime tab tracks PCE, FF, J_{sc} , and V_{oc} over time by performing periodic I-V measurements and analysis. Between I-V measurements, the solar cell can be held at short-circuit, open-circuit, or maximum power.




Figure 9.2. Ossila Solar Cell I-V software: The Lifetime tab.

9.1.3 Stabilised Current

The Stabilised Current tab lets you measure the evolution of the photogenerated current at specific voltages.




Figure 9.3. Ossila Solar Cell I-V software: The Stabilised Current tab.

9.2 Quick Start Guide

1. Start the Ossila Solar Cell I-V software. The window shown in **Figure 9.1** will open.
2. Choose a measurement type as described in **Section 9.1**.
3. Place your sample in the device holder.
4. Place the device holder beneath your solar simulator.
5. Set the appropriate settings in the software (explained in more detail in **Sections 9.4 - 9.8**).
6. Open the shutter of your solar simulator.
7. Click the 'Measure' button.
 - I. For each pixel, measurements are performed using the chosen measurement settings.
 - II. This process is repeated until all pixels have been measured.
8. If automatic saving is turned on, the measurement data and settings will then be saved.

9.3 Shared Software Settings

The settings in these sections are shared between all measurement types.

9.3.1 System Settings

System Address	 COM12
Voltage Source	SMU 1
Current Range	200 mA
Sampling Rate	4096

Figure 9.4. System settings.

(I) System Address

- Select the COM port or IP address of the connected unit you intend to use (USB and Ethernet connection respectively).
 - I. This box will be populated automatically with the addresses of any units connected to the computer.

(II) Voltage Source

- Select which SMU channel of the Source Measure Unit the test board is connected to.
 - I. 'SMU 1' will be automatically selected when pixel switching is set to 'Automated'.

(III) Current Range

- Select the range of currents to be used for the measurement.
 - I. This defines the upper limit and accuracy of current measurements that can be performed by the unit. The values for each range are given in **Table 9.1**.
 - II. Automatic range selection will start on the lowest current range and automatically switch to higher ranges if the current increases above the maximum for a range.

Table 9.1. Maximum current and accuracy for the current ranges of the Ossila Solar Cell I-V Test System.

Maximum Current	Accuracy	Precision	Resolution
±200 mA	±500 µA	10 µA	1 µA
±20 mA	±10 µA	1 µA	100 nA
±2000 µA	±1 µA	100 nA	10 nA
±200 µA	±100 nA	10 nA	1 nA
±20 µA	±10 nA	1 nA	100 pA

(IV) Sampling Rate

- Select the number of samples to be taken for each data point.
 - A higher number of samples per point will improve the accuracy and precision of the measurement. However, this will increase the time taken for the measurement to be performed.

9.3.2 Device Details

The figure displays two side-by-side panels for 'Device Details' settings. The left panel is for 'Automated' switching, and the right panel is for 'Manual' switching.

Automated Panel (Left):

- Pixel Switching:** Automated (dropdown menu)
- Pixels to Measure:** Checkboxes for 1, 2, 3, 4, 5, 6, 7, 8, 9, 10, 11, 12. All are checked.
- Pixel Area (cm²):** 1.0000 (input field)
- Inverted Device:** Radio buttons for Yes and No. 'No' is selected.

Manual Panel (Right):

- Pixel Switching:** Manual (dropdown menu)
- Number of Pixels:** 4 (input field)
- Pixel Area (cm²):** 1.0000 (input field)
- Inverted Device:** Radio buttons for Yes and No. 'No' is selected.

Figure 9.5. Device Details settings for **automated** (left) and **manual** (right) switching. Selecting **Automated – External** gives the same options as manual switching.

(I) Pixel Switching

- Select whether changing the connected pixel is done manually, automatically by the system, or whether to use the external connection on the automated system.

(II) Pixels to Test (Automated Switching)

- Select which pixels to measure.
 - The pixel numbers are labelled on the device holder.

(III) Number of Pixels (Automated – External and Manual Switching)

- Set the number of individual solar cell pixels in the device being measured.

(IV) Pixel Area

- Set the area in cm^2 of each pixel in the device.

(V) Inverted Device

- Set whether the device to be measured is inverted.
 - I. This option should be on if the anode of your device connects to the 'cathode' pins in the device holder.

9.4 Characterisation Settings

9.4.1 Measurement Settings

Start Voltage (V)	-1.00
End Voltage (V)	1.00
Voltage Increment (V)	0.050
Voltage Settle Time (s)	0.000
Illumination ($\text{mW} \cdot \text{cm}^{-2}$)	100.000
Hysteresis I-V	☐ On ☒ Off
Use Shutter	☐ On ☒ Off

Figure 9.6. Measurement Settings for the characterisation and lifetime measurements.

(I) Start Voltage

- Set the voltage in volts at which to start the current-voltage measurement.
 - I. This can be set between -10 V and +10 V.

(II) End Voltage

- Set the voltage in volts at which to end the current-voltage measurement.
 - I. This can be set between -10 V and +10 V.

(III) Voltage Increment

- Set the step size in volts for changing the voltage during current-voltage measurement.

(IV) Voltage Settle Time

- Set the time in seconds between applying a voltage and measuring the current.

- I. This has a maximum of 10 seconds.

(V) Illumination

- Set the illumination intensity (in mW.cm^{-2}) being used during the measurement.

(VI) Hysteresis I-V

- This option performs a reverse current-voltage measurement after the forward current-voltage measurement has completed.
 - I. This reverses the set start and end voltages and uses the same voltage increment and settle time as the forward measurement.

(VII) Use Shutter

- If set to On, the Shutter connection on the Source Measure Unit and automated system will be activated during a measurement and deactivated afterwards.
 - I. When activated, the shutter outputs a constant 5 V signal.

9.5 Lifetime Settings

9.5.1 Measurement Settings

Start Voltage (V)	-1.00
End Voltage (V)	1.00
Voltage Increment (V)	0.050
Voltage Settle Time (s)	0.000
Illumination (mW.cm^{-2})	100.000
Hysteresis I-V	☐ On ☒ Off
Use Shutter	☐ On ☒ Off

Figure 9.7. Measurement Settings for the characterisation and lifetime measurements.

(I) Start Voltage

- Set the voltage in volts at which to start the current-voltage measurement.
 - I. This can be set between -10 V and +10 V.

(II) End Voltage

- Set the voltage in volts at which to end the current-voltage measurement.

- I. This can be set between -10 V and +10 V.

(III) Voltage Increment

- Set the step size in volts for changing the voltage during current-voltage measurement.

(IV) Voltage Settle Time

- Set the time in seconds between applying a voltage and measuring the current.
 - I. This has a maximum of 10 seconds.

(V) Illumination

- Set the illumination intensity in mW.cm^{-2} being used during the measurement.

(VI) Hysteresis I-V

- Set whether to perform a reverse current-voltage measurement after the forward current-voltage measurement has completed.
 - I. This reverses the set 'start' and 'end' voltages and uses the same voltage increment and settle time as the forward measurement.

(VII) Use Shutter

- If set to **On**, the Shutter connection on the Source Measure Unit and automated system will be activated during a measurement and deactivated afterwards.
 - I. When activated, the shutter outputs a constant 5 V signal.

9.5.2 Lifetime Parameters



Duration (Hours)	72.00
Interval (mins)	1.00
Hold Voltage	Maximum Power

Figure 9.8. Lifetime Parameters settings.

(I) Duration

- Set the total duration in hours of the lifetime measurement.

(II) Interval

- Set the time interval in minutes between performing repeat current-voltage measurements of the device.


(III) Hold Voltage

- Set the voltage that all pixels will be held at between measurements.
- This can be set as:
 - I. Short-Circuit – hold at 0 V.
 - II. Maximum Power – hold at the average maximum power point determined from most recent current-voltage curve.
 - III. Open-Circuit – hold at the average open-circuit voltage determined from the most recent current-voltage curve.

Note: As the voltage source is a single channel, the hold voltage will be the same for all pixels being tested.

9.6 Stabilised Current Settings

9.6.1 Measurement Settings



Voltage (V)	1.00
Total Time (mins)	1.00
Measurement Interval (s)	0.00
Use Shutter	☐ On ☒ Off

Figure 9.9. Experimental Parameters settings for the Stabilised Current Output.

(I) Voltage

- Set the voltage to apply to the sample for the measurement.
 - I. This can be set between -10 V and +10 V.

(II) Total Time

- Set the total length of the measurement in minutes.

(III) Measurement Interval

- Set the time between each current measurement in seconds.

(IV) Use Shutter

- If set to **On**, the Shutter connection on the Source Measure Unit and automated system will be activated during a measurement and deactivated afterwards.
 - I. When activated, the shutter outputs a constant 5 V signal.

9.7 Saving and Loading Settings




Figure 9.10. Controls for saving and loading settings profiles.

(I) Save Settings

- Saves the current settings as a profile that can be loaded quickly for use at another time.
- When clicked, you will be prompted to name the settings profile.
 - I. If the name is already in use, you will be asked if you wish to overwrite the previous profile.
 - II. The name cannot contain the characters: \ / : * ? " < > |
- The settings profile will be added to the drop-down box using the given name.

(II) Settings Profiles

- Select a saved settings profile from the drop-down box.
 - I. The settings fields will be populated with the saved values.
- Settings profiles can be deleted by selecting the profile, and then clicking the red 'delete' button next to the drop-down box.

9.8 Saving Results




Figure 9.11. Saving measurement data settings.

(I) Save After Measurement

- Set whether the measurement data will be saved after the measurement has completed.

Warning: Automatic saving can be turned off for lifetime measurements. However, **manual saving is unavailable** for lifetime measurements, so you will not be able to save your data if it is turned off.

- The program allows for data to be saved automatically and manually once the measurement is complete.
 - I. For automatic saving, the 'Saving' fields must be filled in before the measurement can start, these are detailed below.
- For all measurements, a save directory must be specified. This can be done either by:
 - I. Manually typing the directory into the 'Save Directory' field,
 - II. Copy and pasting from your file explorer,
 - III. Clicking the file icon button, which will open a dialog box to allow the selection of a folder to save to.
- All output files are comma separated variable (.csv) files.

(II) Save Directory

- Set the directory in which to create the data files.
- This can be filled in by:
 - I. Manually typing the directory into the 'Save Directory' field,
 - II. Copy and pasting from your file explorer,
 - III. Clicking the 'Select Directory' button, which will open a dialog box to allow the selection of a folder to save to.

(III) Device Label

- Set the name of the device being tested.
- This is used to label the files for I-V data and measurement settings.
- This field cannot contain the following characters:
 - I. \ / : * ? " < > |

(IV) Save Data Format

- All data is saved as .csv (comma separated value) files.
- The figures below show the files that are created when saving data for each of the measurements.




Figure 9.12. Characterisation save data format.




Figure 9.13. Stabilised Current Output save data format.




Figure 9.14. Solar Lifetime Measurement save data.

9.9 Controls



Figure 9.15. Controls for the measurements.

(I) Measure

- Clicking this button will start the measurement using the chosen settings.
- This button cannot be clicked if the software has not detected the test system.

(II) Abort

- Stops a measurement that is currently in progress.

9.10 Plot Controls

(I) Position Readout

Whilst the mouse cursor is over the plot in the Characterisation and Stabilised Current tabs, the x and y position of its location are displayed in the bottom-right of the plot, as shown in Figure 9.16.




Figure 9.16. Readout of the voltage and current density at the mouse cursor location in the Characterisation tab.

(II) Plot Display Controls

By default, the axes of the plot will automatically scale to display all the data within it. The view can be controlled manually using the following mouse controls:

- Left/Middle click and drag – pan the axes.
- Right click and drag – scale the axes (left-right for x-axis, up-down for y-axis).
- Scroll wheel – scale the axes centred on the cursor location.

A specific axis can be controlled by using these controls on the axis labels. The axes can be reset by clicking the 'A' button in the bottom-left of the plot, as shown in Figure 9.17.




Figure 9.17. Button to reset the plot axes.

10. Test Devices

The system is shipped with a test device that can be used to check the calibration of the system. They have resistors arranged in the geometry of the substrate pixels, and the appearance of the test device will depend on the substrate system being used (**Figure 10.1**).




Figure 10.1. Test device configurations.

10.1 Taking a Measurement

1. Plug in and switch on the system.
2. Allow at least 30 minutes for the system to warm up.
3. Place the test device in the device holder with the resistors facing up for S211 and S2006 and facing down for S241/S251.
4. Start the Solar Cell I-V software and enter the following settings in **Figure 10.2**.
 - I. These settings can be used with any current range except for the 20 μA range. For this range the start and end voltages must be lowered to -2 V and 2 V respectively.
 - II. The 'Pixels to Measure' checkboxes (Automated systems) or 'Number of Pixels' (Manual systems) should match the device configuration you have.

Automated Settings:

- Pixel Switching: Automated
- Pixel 1: ☒ Pixel 2: ☒ Pixel 3: ☒ Pixel 4: ☒
- Pixel 5: ☒ Pixel 6: ☒ Pixel 7: ☒ Pixel 8: ☒
- Pixel 9: ☐ Pixel 10: ☐ Pixel 11: ☐ Pixel 12: ☐
- Pixel Area (cm²): 1.0000
- Inverted Device: ☐ Yes ☒ No

Manual Settings:

- Pixel Switching: Manual
- Number of Pixels: 4
- Pixel Area (cm²): 1.0000
- Inverted Device: ☐ Yes ☒ No

Measurement Parameters:

- Start Voltage (V): -10.00
- End Voltage (V): 10.00
- Voltage Increment (V): 1.000
- Voltage Settle Time (s): 0.000
- Illumination (mW.cm⁻²): 100.000
- Hysteresis I-V: ☐ On ☒ Off

Figure 10.2. Measurement settings for the resistor test device.

5. Click the 'Measure' button.
6. The system should measure straight line resistor responses from $-100 \mu\text{A.cm}^{-2}$ to $100 \mu\text{A.cm}^{-2}$ (or $-20 \mu\text{A.cm}^{-2}$ to $20 \mu\text{A.cm}^{-2}$ for the $20 \mu\text{A}$ current range) as shown in **Figure 10.3**.
7. To check the calibration of the system, use the I-V data to calculate the measured resistance at -10 and 10 V (-2 and 2 V for the $20 \mu\text{A}$ current range).
 - I. Resistance can be calculated using: $R = V / I$
 - II. For the 200 mA current range the calculated resistance should be between 98 and 102 k Ω (within 2% of the resistor value).
 - III. For all other ranges the calculated resistance should be between 99 and 101 k Ω (within 1% of the resistor value).




Figure 10.3. Example measurement of resistor test device using the 200 μA current range.

11. Troubleshooting

Most of the issues that may arise will be detailed here. However, if you encounter any issues that aren't in this list, please contact us by email at info@ossila.com, and we will respond as soon as possible.

11.1 Installation and Setup

Problem	Possible Cause	Action
No power	The power supply may not be connected properly.	Ensure the system is firmly plugged into the power supply, and that the plug is connected to both the adaptor and a working power socket.
	The power supply adaptor has a fault.	Contact Ossila for a replacement power supply adaptor.
Software does not start	The wrong version of Windows is installed on the computer.	Install the software on a computer with Windows 10 or newer.
	The software has not installed properly.	Try reinstalling the software.
Cannot connect to the system via USB	The USB cable may not be connected properly.	Ensure the USB cable is firmly plugged in at both ends.
	The USB cable may not be connected to a working USB port.	Try connecting the unit to a different USB port on the computer.
	The USB cable is defective.	Try using a different USB-B cable, and contact Ossila if necessary.
Cannot connect to the system via network	The MAC address of the unit is not registered with the internal network.	Register the system on the network using the MAC address obtained via a USB connection (see Source Measure Unit manual).
	The Ethernet cable may not be connected properly.	Ensure the Ethernet cable is firmly plugged in at both ends.
	The Ethernet cable is defective.	Try using a different Ethernet cable.

11.2 Error Messages and Warnings

Message	Description
Current compliance reached	The measured current is greater than the set current limit.
Error communicating with system	The software is unable to connect to the system.
No device holder detected	The device holder is not connected to the system properly.
Voltage increment cannot be zero	The voltage increment is set to 0 V.
Start and end voltage cannot be equal	The start and end voltages of are set to the same voltage.
No save directory or device label entered	The save directory and/or the device label fields are empty.
Settings profile not found	The given settings profile does not exist or is open in other software.
Error loading settings	There is a problem with the settings profile preventing it from being loaded.
Error deleting profile	The given settings profile does not exist, the software does not have the necessary permissions to delete it, or it is open in other software.
No data to save	There is no measurement data in memory to save to file.
Error saving data	The software does not have the necessary permissions to access the given file path, or the file is already open in other software.
Error saving settings	The software does not have the necessary permissions to access the given file path, or the file is already open in other software.
Error creating data directories	The software could not create the directories for save data files.
Error creating data files	The software could not create the files for saving measurement data.

12. Related Products

12.1 Related Consumables



ITO Coated Substrates

Our range of ITO substrates for OPV, OLED, and sensing applications.

Product codes: S111 / S101 / S211 / S281 / S171



Flat Tip Tweezers

Provides a good substrate grip without scratching.

Product code: C121



FTO Coated Substrates

Designed to be used as transparent electrodes for thin-film photovoltaics.

Product codes: S301 / S302 / S303 / S304



Substrate Cleaning Rack

Holds 20 substrates for a variety of processing techniques.

Product code: E101

12.2 Related Equipment



Spin Coater

Product high-quality coatings without any substrate warping. Perfect for busy labs with limited space.

Product code: L2001A3



Solar Simulator

Compact, low price light source suitable for characterising small area solar cells.

Product codes: G2009A1 / G2009B1



UV Ozone Cleaner

For removing contamination on the surface of samples, providing you with ultraclean surfaces within minutes.

Product code: L2002A2



Source Measure Unit

Source voltage, measure current, get data. Simplify and accelerate your data collection!

Product code: P2005A2